



INTERNATIONAL QUALITY AND ACCREDITATION SERVICES LLP

Scope of Accreditation for Calibration

As per ISO/IEC 17025:2017

Page 1 of 1

CAB Name : Plausible Instruments Pvt. Ltd.
Address : 1/5241-A, St. No. 8, Balbir Nagar, Shahdara, Delhi 110032

Certificate No: C-0001

Issue date : 13.10.2023
Valid till : 12.10.2024
Amendment date : NA

Mechanical

Sr. No.	Parameter/Measurand quantity, Instrument or gauge	Range	CMC(±) @ k=2, approx. 95% CF	Remarks/Method used	Facility
01	Test Sieve	4.75 to 125 mm	27.5 µm	Comparison	Permanent
02	Elongation Gauge	Upto 81 mm	47 µm	Comparison	Permanent
03	Flakiness Gauge	Upto 100 mm	50 µm	Comparison	Permanent

Dr. Himani Jain
Founder, IQAS

Ratan Singh Rana
Founder, IQAS